



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE

United States Patent and Trademark Office

Address: COMMISSIONER FOR PATENTS

P.O. Box 1450

Alexandria, Virginia 22313-1450

www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/582,914	06/14/2006	Takenari Ito	1027550-000188	8662
21839 7590 03/31/2010 BUCHANAN, INGERSOLL & ROONEY PC POST OFFICE BOX 1404 ALEXANDRIA, VA 22313-1404				
EXAMINER				
GRAY, PHILLIP A				
ART UNIT		PAPER NUMBER		
3767				
NOTIFICATION DATE		DELIVERY MODE		
03/31/2010		ELECTRONIC		

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

Notice of the Office communication was sent electronically on above-indicated "Notification Date" to the following e-mail address(es):

ADIPFDD@bipc.com  
offserv@bipc.com

### Office Action Summary

**Application No.**

10/582,914

**Applicant(s)**

ITOU ET AL.

**Examiner**

Phillip Gray

**Art Unit**

3767

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --  
**Period for Reply**

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

**Status**

- 1) ☒ Responsive to communication(s) filed on 15 February 2010.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

**Disposition of Claims**

- 4) ☒ Claim(s) 1-20 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-20 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

**Application Papers**

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

**Priority under 35 U.S.C. § 119**

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

**Attachment(s)**

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO/SE/US)  
Paper No(s)/Mail Date \_\_\_\_\_
- 4) ☐ Interview Summary (PTO-413)  
Paper No(s)/Mail Date \_\_\_\_\_
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: \_\_\_\_\_

**DETAILED ACTION**

This office action is in response to applicant's communication of 2/15/2010.  
Currently amended claims 1-20 are pending and rejected below.

***Claim Rejections - 35 USC § 103***

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

This application currently names joint inventors. In considering patentability of the claims under 35 U.S.C. 103(a), the examiner presumes that the subject matter of the various claims was commonly owned at the time any inventions covered therein were made absent any evidence to the contrary. Applicant is advised of the obligation under 37 CFR 1.56 to point out the inventor and invention dates of each claim that was not commonly owned at the time a later invention was made in order for the examiner to consider the applicability of 35 U.S.C. 103(c) and potential 35 U.S.C. 102(e), (f) or (g) prior art under 35 U.S.C. 103(a).

Claims 1-20 are rejected under 35 U.S.C. 103(a) as being unpatentable over Cook (U.S. Patent 4,637,396) in view of Fina (U.S. Patent 4,911,163).

Cook discloses a balloon catheter with an outer catheter with and inner layer (22), outer layer (24) and reinforcing layer (23) and outer catheter hub (46) at the

proximal end, an inner catheter (13) with a hard proximal part (40) softer distal part (18), and inner catheter hub (55), and second soft tip (17) on the proximal end (see figure 1-3) wherein said outer catheter hub and inner hub are adapted to be fixed to each other as not to rotate and when engaged the inner catheter protrudes from the outer catheter with the distance no more than 10 mm (see figure 1). Concerning claim 3, 8, and 18-20 note lower portion of figure 2. Concerning claim 4 and 9-10 note portion near 28-29. Concerning claim 5 note figure 1. Concerning claim 6 and 11-14 see layer 19.

Concerning the newly amended claim language of the outer catheter hub and said inner catheter hub are disengaged from each other so that said inner catheter is removable from said outer catheter, it is examiners position that the prior art reference would have the ability to disengage from each other, if joints at 53 or 46 in figure 5 where unscrewed or displaced then the outer and inner catheter/hubs would be removable and disengaged from one another. Further it is examiners position that figure 5 and figure 1 shown the device when the outer and inner hubs are engaged with each other and the two catheters do not move relative to one another.

Cook discloses the claimed invention except for the locking and removable inner catheter. Fina teaches that it is known to use a removable inner catheter which also locks as set forth in paragraphs at column 3-6 discussing how catheter 2 relates to catheter 8, to provide two catheters that may function jointly or independently of one another. It would have been obvious to one having ordinary skill in the art at the time the invention was made to modify the system as taught by Cook with a removable inner catheter as taught by Fina, since such a modification would provide the System with a

removable inner catheter for providing two catheters that may function jointly or independently of one another.

Claims 7, 15-17 are rejected under 35 U.S.C. 103(a) as being unpatentable over Cook in view of Fina. Cook in view of Fina discloses the claimed invention except for the .90 ratio of the outside to inside catheter diameters. It would have been obvious to one having ordinary skill in the art at the time the invention was made to construct a .90 ratio of the outside to inside catheter diameters., since such a modification would have involved a mere change in the size of a component. A change in size is generally recognized as being within the level of ordinary skill in the art. *In re Rose*, 105 USPQ 237 (CCPA 1955).

### ***Response to Arguments***

Applicant's arguments filed 7/6/2009 have been fully considered but they are not persuasive. Concerning the amendment to claim 1, of the outer and inner hub can be disengaged from each other so that said inner catheter is removable from said outer catheter, it is examiner position that the prior art discloses and would have the ability to perform this function. Further more it has been held that constructing a formerly integral structure in various elements involves only routine skill in the art. Therefore it is examiners position that making a inner catheter removable from an outer catheter is an obvious modification. See rejection above.

### ***Conclusion***

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Phillip Gray whose telephone number is (571)272-7180. The examiner can normally be reached on Monday through Friday, 8:30 a.m. to 4:30 p.m. EST.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Kevin Simons can be reached on (571) 272-4965. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Phillip Gray/  
Examiner, Art Unit 3767  
/Kevin C. Simons/  
Supervisory Patent Examiner, Art Unit 3767